| Clinical Situation  | <b>Appliance Choices</b>   | Comments  |
|---|--|---|
| Medicare patient  Some BCBS (Blue Cross Blue Shield) and UnitedHealthcare plans | Medicare TAP and Herbst approved appliances  | 1. In selection for our elderly patients, consider the ease of usage and adjustability if you allow them to perform titration at home.        |
|   |  | 2. Some medical plans require a PDAC (Pricing, Data Analysis and Coding)-approved appliance for use for their insured.                        |
| Buccal exostosis (maxilla)  | Any device that does not require a mechanism to be on the buccal aspect that may approximate the tori. | Verify with lab for a specific appliance design that would not interfere with the exostosis.  |
| Edentulous  | 1. Tongue retaining device (TRD)   | 1. For those with poor ridges where implants are unable to be placed, a TRD may be appropriate.   |
|   | 2. Implant-retained appliances   | 2. With an implant-retained appliance, it is preferable to use an acrylic-fabricated device.  |
|   | 3. If dental ridges are adequate, place a traction-based appliance over the denture                    | 3. A chinstrap may be required to allow approximation of the dental ridges to engage into the denture, which would hold the mandible forward. |
| Partial edentulous  | Any appliance  | Verify which teeth need to be present for a specific appliance design.  |
| Short clinical crowns or poor retention   | Acrylic-based appliances   | Ball clasps to assist with retention may be required.   |
| Open bite   | Narval CC<br>Panthera D-SAD  | Excess vertical is not required for strength on the posterior aspect of these appliances.   |
| Anterior mobile teeth   | Narval CC<br>Panthera D-SAD  | Design has no contact of the buccal strap over the anterior incisor.  |

| Bruxer                         | The Silencer       | If clinical situation permits, |
|--------------------------------|--------------------|--------------------------------|
|                                | Herbst             | the Narval CC and Panthera     |
|                                | Narval CC          | D-SAD are very durable.        |
|                                | Panthera D-SAD     |                                |
| Material sensitivity (acrylic) | Narval CC          | These are fabricated from      |
|                                | Panthera D-SAD     | an organic biocompatible       |
|                                |                    | polymer material.              |
| Compliance tracking            | SomnoMed series    | As of this press time, the     |
|                                | appliances and TAP | Braebon DentiTrac is           |
|                                | appliances         | available with these           |
|                                |                    | manufacturers.                 |

Content provided by Mayoor Patel, DDS, MS, and published by *Sleep Review*. Cannot be reprinted or distributed without permission. Dentists can subscribe to *Sleep Review* for free at www.sleepreviewmag.com/subscribe.